



Member of the FM Global Group

FM Approvals  
1151 Boston Providence Turnpike  
P.O. Box 9102 Norwood, MA 02062 USA  
T: 781 762 4300 F: 781-762-9375 www.fmaprovals.com

# CERTIFICATE OF COMPLIANCE

## HAZARDOUS LOCATION ELECTRICAL EQUIPMENT PER CANADIAN REQUIREMENTS

This certificate is issued for the following equipment:

***XMT9ab-1-1-c-d-e. Digital Flow Transmitter.***

XP/I/1/BCD/T6 Ta = -40°C to +60°C; IP67

a = Branding name and software type 0, 1, 2, 3, 4, 5, 6, 7, 8 or 9.

b = Branding name and software type 0, 1, 2, 3, 4, 5, 6, 7, 8 or 9.

c = Power 1 or 2.

d = Display 0 or 1.

e = I/O card A or B.

***XMT9abJB-1-1. Junction Box.***

XP/I/1/BCD/Ta = -40°C to +105°C; IP67

a = Branding name 0, 1, 2, 3, 4, 5, 6, 7, 8 or 9.

b = Branding name 0, 1, 2, 3, 4, 5, 6, 7, 8 or 9.

Equipment Ratings:

Explosionproof for Class I, Division 1, Groups B, C and D hazardous locations, indoors and outdoors (IP67).

FM Approved for:

GE Infrastructure Sensing  
Billerica, Massachusetts, USA

To verify the availability of the Approved product, please refer to [www.approvalguide.com](http://www.approvalguide.com)



This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

C22.2 No. 0.4	2009
C22.2 No. 0.5	2008
C22.2 No. 30	2007
C22.2 No. 1010.1	2009
C22.2 No. 60529	2010

Original Project ID: 3047150C

Approval Granted: July 18, 2013

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Approvals LLC

  
\_\_\_\_\_  
J. E. Marquedant  
Group Manager, Electrical

18 July 2013  
\_\_\_\_\_  
Date

To verify the availability of the Approved product, please refer to [www.approvalguide.com](http://www.approvalguide.com)